

Certificate of Compliance

Product : LCD Monitor
Applicant : AG Neovo
Rating(s) : Power Adapter: Adapter Technology Co., Ltd. /
ATM065T-P240
Input: 100-240Vac, 50-60Hz, 1.6-0.7A;
Output: 24Vdc, 2.71A

Model/Type : DR-17Gx, DR-22Gx (where x can be blank, A to Z (a-z), 0 to
9 or "+", "-", "\", "/")
LCD Monitor rating: 24Vdc, 2.71A

Test Report No. : M190522-02-A0

Standard(s) Applied : EN 60601-1:2006 + A11: 2011 + A1: 2013 + A12: 2014;
IEC 60601 1:2005 (Third Edition) + CORR. 1:2006 + CORR.
2:2007 + A1:2012
(or IEC 60601-1: 2012 reprint)

We hereby declare that the above equipment has been tested in our facility with the listed standard and found in compliance with the harmonized Standards of Medical Device Directive 93/42/EEC



A handwritten signature in black ink, appearing to read 'Tony Hsu', is written over a horizontal line.

Tony Hsu / Manager

2019-08-29



Prodigy Technology Consultant Co.Ltd.

No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan CHINESE TAIPEI
TEL: 886-2-2603-7288/FAX: 886-2-2602-0908