


# Certificate of Compliance

Product : LCD Monitor  
Applicant : Associated Industries China, Inc.  
Rating(s) : 24Vdc, 1.5A  
Model/Type : DR-22E  
Test Report No. : M170811-04-A0  
Standard(s) Applied : EN 60601-1:2006 + A11: 2011 + A1: 2013  
+ A12: 2014;  
IEC 60601-1: 2005 + CORR. 1:2006 +  
CORR. 2:2007 + AM1:2012

We hereby declare that the above equipment has been tested in our facility with the listed standard and found in compliance with the harmonized Standards of Medical Device Directive 93/42/EEC

**CE**



Tony Hsu / Senior Engineer

2017-11-17



**Prodigy Technology Consultant Co.Ltd.**

No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan CHINESE TAIPEI  
TEL: 886-2-2603-7288/FAX: 886-2-2602-0908