


Certificate of Compliance

Product : LCD Monitor

Applicant : AG Neovo

Rating(s) : Power Adapter: Adapter Technology Co., Ltd. /
ATM065T-P120
Input: 100-240Vac, 50-60Hz, 1.6-0.7A;
Output: 12Vdc, 5.0A
LCD Monitor rating: 12Vdc, 2.5A

Model/Type : MX-24x, MX-2402, MX-24xxxx (where x can be A to Z
(a-z), 0 to 9 or "+", "-", "\", "/")


Trademark : 

Test Report No. : M240520-06-3X-A0

Standard(s) : EN 60601-1:2006/A2:2021;
IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC

Applied : 60601-1:2005/AMD2:2020

We hereby declare that the above equipment has been tested in our facility with the listed standard and found in compliance with the harmonized Standards of Medical Device Regulation 2017/745



Tony Hsu / Manager

2024-06-27



Prodigy Technology Consultant Co.Ltd.

No. 12, Gong 7th Rd., Linkou District, New Taipei City 24450, Taiwan Chinese Taipei
TEL: 886-2-2603-7288/FAX: 886-2-2602-0908