

CERTIFICATE OF CONFORMITY



BUREAU
VERITAS

UK
CA

Product : LCD monitor
Brand : AG neovo
Model No. : MX-2402
Series Model : MX-24xxxxx (The "x" can be any alphanumeric character including blank, for marketing difference)
Applicant : Top Victory Electronics (Taiwan) Co., Ltd.
Report No. : CEBEFM-WTW-P24060116

We, **Bureau Veritas Consumer Products Services (H.K.) Ltd., Taoyuan Branch**, declare that the equipment above has been tested in our facility and found compliance with the requirement limits of applicable standards, in accordance with the Medical Devices Regulations 2002 (S.I. 2002/618). The test record, data evaluation and Equipment Under Test (EUT) configurations represented herein are true and accurate under the standards herein specified.

EN 60601-1-2: 2015+A1:2021

EN 60601-1-2: 2015

IEC 60601-1-2: 2020 ED. 4.1

IEC 60601-1-2: 2014 ED. 4.0

EN 55032:2015 +A11:2020 / CISPR 32:2015 +Cor 1:2016, Class B

EN 61000-3-2: 2014, Class A

EN IEC 61000-3-2: 2019+A1:2021, Class A

EN 61000-3-3:2013 +A1:2019 +A2:2021

EN 61000-4-2: 2009 / IEC 61000-4-2: 2008 ED. 2.0

EN IEC 61000-4-3: 2020 / IEC 61000-4-3: 2020 ED. 4.0

EN 61000-4-4: 2012 / IEC 61000-4-4: 2012 ED. 3.0

EN 61000-4-5: 2014+A1:2017 / IEC 61000-4-5: 2017 ED. 3.1

EN 61000-4-6: 2014+AC:2015 / IEC 61000-4-6: 2013 ED. 4.0

EN 61000-4-8: 2010 / IEC 61000-4-8: 2009 ED. 2.0

EN IEC 61000-4-11: 2020+AC:2020 / IEC 61000-4-11: 2020 ED. 3.0

EN 61000-4-39: 2017 / IEC 61000-4-39: 2017 ED. 1.0

NOTE: The above EN/IEC basic standards are applied with latest version if customer has no special requirement.

Jim Hsiang / Associate Technical Manager

2024/7/23



No. 47-2, 14th Ling, Chia Pau Vil., Lin Kou Dist., New Taipei City, Taiwan

Tel: 886-2-26052180 Fax: 886-2-26051924

<http://ee.bureauveritas.com.tw> E-Mail: service.adt@tw.bureauveritas.com